

Monthly radioactivity measurement "June, 2011" was conducted.**Measurement Result:**

No radioactive substances (Iodine/Caesium) was detected at either Unisis and subcontractors, with Products/Cleaning Water/Cannulae/Components.

Whilst, Caesium was detected with Air-Conditioner's Filter at external-air intake. Please re

Radioactivity measurement - June/2011

Sample Category	Nuclide					
	Iodine	Caesium			Radiation Dose	
	I-131	Cs-134	Cs-136	Cs-137		
Unit	Bq/kg(L)	Bq/kg(L)	Bq/kg(L)	Bq/kg(L)	μSv/h	cpm
Products (Needles)	ND ^{*1}	ND ^{*1}	ND ^{*1}	ND ^{*1}	-	-
Cleaning Water	ND ^{*1}	ND ^{*1}	ND ^{*1}	ND ^{*1}	-	-
Cannulae	—	—	-	—	ND ^{*2}	-
Components	—	—	-	—	ND ^{*2}	-

*1) Not Detected (Detection Limit=6~8)

*2) Not Detected (Equivalent count rate as Background)

Air Filter for internal circulation is scheduled to be measured at the beginning of July.

For detail please refer to May Report.

END